New Bloom Labs
6121 Heritage Park Dr.
Chattanooga, TN 37416

(844) 837-8223
http://newbloomlabs.com
ISO/IEC 17025:2017 Testing Laboratory Cert# AT-2868

Stay Cool Beverages - Berry Delight - Delta-9

Sample ID: 2605NBL0894.2468
Matrix: Ingestible
Type: Beverage
Sample Size:
Date Collected:

Received: 05/22/2026
Completed: 06/09/2026
Expires: 06/09/2027
External Lot ID: 5/8/2028
Batch#: 6019

Client
Stay Cool
Lic. #
2800 Treble Lane #834 Austin, TX 78704
(512) 971-1316 steven@drinkstaycool.com

Summary

Test	Date Tested	Result
Cannabinoids	05/23/2026	Complete
Residual Solvents	06/04/2026	Complete
Microbials	06/01/2026	Complete
Mycotoxins	06/05/2026	Complete
Pesticides	06/05/2026	Complete
Heavy Metals	06/09/2026	Complete
Foreign Matter	05/22/2026	Not Detected

Cannabinoids

Complete

9.807 mg/unit	0.0027%	ND	9.807 mg/unit
Total THC	Δ9-THC	Total CBD	Total Cannabinoids

Analyte	LOD	LOQ	Result	Result	Result
	mg/unit	mg/unit	mg/unit	mg/mL	%
(6aR,9R)-d10-THC	0.9133	1.370	ND	ND	ND
9R-HHC	0.9133	1.370	ND	ND	ND
(6aR,9S)-d10-THC	0.9133	1.370	ND	ND	ND
9S-HHC	0.9133	1.370	ND	ND	ND
CBC	0.9133	1.370	ND	ND	ND
CBCa	0.9133	1.370	ND	ND	ND
CBD	0.9133	1.370	ND	ND	ND
CBDa	0.9133	1.370	ND	ND	ND
CBDV	0.9133	1.370	ND	ND	ND
CBDVa	0.9133	1.370	ND	ND	ND
CBG	0.9133	1.370	ND	ND	ND
CBGa	0.9133	1.370	ND	ND	ND
CBN	0.9133	1.370	ND	ND	ND
CBNa	0.9133	1.370	ND	ND	ND
Δ8-THC	0.9133	1.370	ND	ND	ND
Δ9-THC	0.9133	1.370	9.807	0.02770	0.003
THCa	0.9133	1.370	ND	ND	ND
THCp	0.9133	1.370	ND	ND	ND
THCV	0.9133	1.370	ND	ND	ND
THCVa	0.9133	1.370	ND	ND	ND
Total THC			**9.807**	**0.02770**	**0.003**
Total CBD			**ND**	**ND**	**ND**
Total			**9.807**	**0.02770**	**0.003**

Date Tested: 05/23/2026
Density: 1 mL = 1.032 g, Unit Mass: 365.328 g, 1 Unit = 12 fl oz
Testing Method: HPLC-UV, CON-P-3000; Validation Date: 05/2019.
Total THC = THCa * 0.877 + Δ9-THC; Total CBD = CBDa * 0.877 + CBD; LOQ = Limit of Quantitation; LOD = Limit of Detection; ND = Not Detected; Total THC Measurement of Uncertainty: ± 0.040%, Total CBD Measurement of Uncertainty: ± 2.000%.

Ashley Phillips
Laboratory Director
06/09/2026

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com

All analyses were conducted at 6121 Heritage Park Dr, Suite A500 Chattanooga, TN 37416. Results published on this certificate relate only to the items tested. Items are tested as received. New Bloom Labs makes no claims as to the efficacy, safety, or other risks associated with any detected or non-detected level of any compounds reported herein. This certificate shall be reproduced in full, except with the written approval of New Bloom Labs. Measurement uncertainties are determined in accordance with ISO 17025 and are based on the total expanded uncertainty with a 95% confidence interval (k=2). Filth and Foreign Testing Method - CON-P-11000.

Stay Cool Beverages - Berry Delight - Delta-9

Sample ID: 2605NBL0894.2468
Matrix: Ingestible
Type: Beverage
Sample Size:
Date Collected:

Received: 05/22/2026
Completed: 06/09/2026
Expires: 06/09/2027
External Lot ID: 5/8/2028
Batch#: 6019

Client
Stay Cool
Lic. #
2800 Treble Lane #834 Austin, TX 78704
(512) 971-1316 steven@drinkstaycool.com

Pesticides

Complete

Analyte	LOQ	LOD	Limit	Result	Status
	PPM	PPM	PPM	PPM	
Abamectin	0.089	0.026		ND	Tested
Acephate	0.085	0.026		ND	Tested
Acequinocyl	0.085	0.026		ND	Tested
Acetamiprid	0.085	0.026		ND	Tested
Aldicarb	0.085	0.026		ND	Tested
Azoxystrobin	0.085	0.026		ND	Tested
Bifenazate	0.085	0.026		ND	Tested
Bifenthrin	0.085	0.026		ND	Tested
Boscalid	0.085	0.026		ND	Tested
Captan	0.342	0.103		ND	Tested
Carbaryl	0.085	0.026		ND	Tested
Carbofuran	0.085	0.026		ND	Tested
Chlorantraniliprole	0.085	0.026		ND	Tested
Chlordane	0.085	0.026		ND	Tested
Chlorfenapyr	0.068	0.021		ND	Tested
Chlormequat	0.085	0.026		ND	Tested
Chlorpyrifos	0.085	0.026		ND	Tested
Clofentezine	0.085	0.026		ND	Tested
Coumaphos	0.085	0.026		ND	Tested
Cyfluthrin	0.427	0.128		ND	Tested
Cypermethrin	0.427	0.128		ND	Tested
Daminozide	0.855	0.256		ND	Tested
Diazinon	0.085	0.026		ND	Tested
Dichlorvos	0.043	0.014		ND	Tested
Dimethoate	0.085	0.026		ND	Tested
Dimethomorph	0.082	0.026		ND	Tested
Ethoprophos	0.085	0.026		ND	Tested
Etofenprox	0.085	0.026		ND	Tested
Etoxazole	0.085	0.026		ND	Tested
Fenhexamid	0.085	0.026		ND	Tested
Fenoxycarb	0.085	0.026		ND	Tested
Fenpyroximate	0.085	0.026		ND	Tested
Fipronil	0.085	0.026		ND	Tested
Flonicamid	0.085	0.026		ND	Tested
Fludioxonil	0.085	0.026		ND	Tested
Hexythiazox	0.085	0.026		ND	Tested
Imazalil	0.085	0.026		ND	Tested
Imidacloprid	0.085	0.026		ND	Tested
Kresoxim Methyl	0.085	0.026		ND	Tested
Malathion	0.085	0.026		ND	Tested
Metalaxyl	0.085	0.026		ND	Tested
Methiocarb	0.085	0.026		ND	Tested
Methomyl	0.085	0.026		ND	Tested
Mevinphos	0.085	0.026		ND	Tested
Myclobutanil	0.085	0.026		ND	Tested
Naled	0.214	0.065		ND	Tested
Oxamyl	0.085	0.026		ND	Tested
Paclobutrazol	0.085	0.026		ND	Tested
Methyl Parathion	0.068	0.021		ND	Tested
Pentachloronitrobenzene	0.068	0.021		ND	Tested
Permethrins	0.085	0.026		ND	Tested
Phosmet	0.085	0.026		ND	Tested
Piperonyl Butoxide	0.085	0.026		ND	Tested
Prallethrin	0.085	0.026		ND	Tested
Propiconazole	0.085	0.026		ND	Tested
Propoxur	0.085	0.026		ND	Tested
Pyrethrins	0.427	0.128		ND	Tested
Pyridaben	0.085	0.026		ND	Tested
Spinetoram	0.085	0.026		ND	Tested
Spinosad	0.087	0.026		ND	Tested
Spiromesifen	0.085	0.026		ND	Tested
Spirotetramat	0.085	0.026		ND	Tested
Spiroxamine	0.085	0.026		ND	Tested
Tebuconazole	0.085	0.026		ND	Tested
Thiacloprid	0.085	0.026		ND	Tested
Thiamethoxam	0.085	0.026		ND	Tested
Trifloxystrobin	0.085	0.026		ND	Tested

Date Tested: 06/05/2026
Testing Method: LC/MS/MS, CON-P-5000; Validation Date: 04/2020. Testing Method: GC/MS/MS, CON-P-12000; Validation Date: 05/2025.

Ashley Phillips
Laboratory Director
06/09/2026

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com

All analyses were conducted at 6121 Heritage Park Dr, Suite A500 Chattanooga, TN 37416. Results published on this certificate relate only to the items tested. Items are tested as received. New Bloom Labs makes no claims as to the efficacy, safety, or other risks associated with any detected or non-detected level of any compounds reported herein. This certificate shall be reproduced in full, except with the written approval of New Bloom Labs. Measurement uncertainties are determined in accordance with ISO 17025 and are based on the total expanded uncertainty with a 95% confidence interval (k=2). Filth and Foreign Testing Method - CON-P-11000.

Stay Cool Beverages - Berry Delight - Delta-9

Sample ID: 2605NBL0894.2468
Matrix: Ingestible
Type: Beverage
Sample Size:
Date Collected:

Received: 05/22/2026
Completed: 06/09/2026
Expires: 06/09/2027
External Lot ID: 5/8/2028
Batch#: 6019

Client
Stay Cool
Lic. #
2800 Treble Lane #834 Austin, TX 78704
(512) 971-1316 steven@drinkstaycool.com

Microbials

Complete

Analyte	LOD	Result	Status
	CFU/g		
Shiga Toxin-producing E. Coli	1.000	ND	Tested
Salmonella	1.000	ND	Tested
Aspergillus spp. (A. flavus, A. fumigatus, A. terreus, A. niger)	1.000	ND	Tested

Testing Method: qPCR, CON-P-6000; Validation Date: 01/2021.

Analyte	LOD	Result	Status
	CFU/g	CFU/g	
Aerobic Bacteria	10	ND	Tested
Enterobacteriaceae (BTGN)	10	ND	Tested
Coliforms	10	ND	Tested
Yeast & Mold	9	9	Tested
E. Coli	10	ND	Tested

Testing Method: Petrifilms, CON-P-9000; Validation Date: 12/2021.

Date Tested: 06/01/2026
CFU = Colony Forming Units, Cq = Quantification Cycle, ND = Not Detected, LOD = Limit of Detection, ND = Not Detected

Mycotoxins

Complete

Analyte	LOD	LOQ	Limit	Result	Status
	PPB	PPB	PPB	PPB	
B1	1	4		ND	Tested
B2	1	4		ND	Tested
G1	1	4		ND	Tested
G2	1	4		ND	Tested
Total Aflatoxins	1	4		ND	Tested
Ochratoxin A	5	17		ND	Tested

Date Tested: 06/05/2026
Testing Method: LC/MS/MS, CON-P-5000; Validation Date: 04/2020.

Heavy Metals

Complete

Analyte	LOD	LOQ	Limit	Result	Status
	PPM	PPM	PPM	PPM	
Arsenic	0.050	0.100		ND	Tested
Cadmium	0.050	0.100		ND	Tested
Lead	0.050	0.100		ND	Tested
Mercury	0.050	0.100		ND	Tested
Palladium	0.050	0.250		ND	Tested
Selenium	0.050	0.100		ND	Tested

Date Tested: 06/09/2026
Testing Method: ICP/MS, CON-P-7000; Validation Date: 09/2020.

Ashley Phillips
Laboratory Director
06/09/2026

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com

All analyses were conducted at 6121 Heritage Park Dr, Suite A500 Chattanooga, TN 37416. Results published on this certificate relate only to the items tested. Items are tested as received. New Bloom Labs makes no claims as to the efficacy, safety, or other risks associated with any detected or non-detected level of any compounds reported herein. This certificate shall be reproduced in full, except with the written approval of New Bloom Labs. Measurement uncertainties are determined in accordance with ISO 17025 and are based on the total expanded uncertainty with a 95% confidence interval (k=2). Filth and Foreign Testing Method - CON-P-11000.

Stay Cool Beverages - Berry Delight - Delta-9

Sample ID: 2605NBL0894.2468
Matrix: Ingestible
Type: Beverage
Sample Size:
Date Collected:

Received: 05/22/2026
Completed: 06/09/2026
Expires: 06/09/2027
External Lot ID: 5/8/2028
Batch#: 6019

Client
Stay Cool
Lic. #
2800 Treble Lane #834 Austin, TX 78704
(512) 971-1316 steven@drinkstaycool.com

Residual Solvents

Complete

Analyte	LOD	LOQ	Limit	Result	Status
	PPM	PPM	PPM	PPM	
Acetone	75.000	250.000		ND	Tested
Benzene	0.300	1.000		ND	Tested
n-Butane	69.000	230.000		ND	Tested
Isobutane	69.000	230.000		ND	Tested
Ethanol	75.000	250.000		537.690	Tested
Ethyl Acetate	75.000	250.000		ND	Tested
Heptane	75.000	250.000		ND	Tested
n-Hexane	1.500	5.000		ND	Tested
2,2-Dimethylbutane	1.500	5.000		ND	Tested
3-Methylpentane	1.500	5.000		ND	Tested
2-Methylpentane/2,3-Dimethylbutane	3.000	10.000		ND	Tested
Methanol	30.000	100.000		ND	Tested
Isopentane	69.000	230.000		ND	Tested
n-Pentane	30.000	100.000		ND	Tested
Neopentane	69.000	230.000		ND	Tested
2-Propanol	75.000	250.000		ND	Tested
Propane	300.000	1000.000		ND	Tested
Toluene	0.150	0.500		ND	Tested
Ethylbenzene	0.150	0.500		ND	Tested
m/p-Xylene	0.150	0.500		ND	Tested
o-Xylene	0.150	0.500		ND	Tested
1,1-Dichloroethene	0.600	2.000		ND	Tested
1,2-Dichloroethane	0.600	2.000		ND	Tested
Ethylene Oxide	1.500	5.000		ND	Tested
Acetonitrile	6.000	20.000		ND	Tested
Chloroform	0.600	2.000		ND	Tested
Ethyl Ether	75.000	250.000		ND	Tested
Dichloromethane	30.000	100.000		ND	Tested
Trichloroethene	6.000	20.000		ND	Tested
2,3-Butanedione	30.000	100.000		ND	Tested

Date Tested: 06/04/2026
Testing Method: HS-GC/MS, CON-P-8000; Validation Date: 09/2020.

Ashley Phillips
Laboratory Director
06/09/2026

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com

All analyses were conducted at 6121 Heritage Park Dr, Suite A500 Chattanooga, TN 37416. Results published on this certificate relate only to the items tested. Items are tested as received. New Bloom Labs makes no claims as to the efficacy, safety, or other risks associated with any detected or non-detected level of any compounds reported herein. This certificate shall be reproduced in full, except with the written approval of New Bloom Labs. Measurement uncertainties are determined in accordance with ISO 17025 and are based on the total expanded uncertainty with a 95% confidence interval (k=2). Filth and Foreign Testing Method - CON-P-11000.